



校准证书  
CALIBRATION CERTIFICATE



证书编号:

Certificate No.



扫一扫验真伪

2WB23031752543-0007

委托方:

Client

陕西高研智能科技有限公司

委托方地址:

Address

西安市高新区西沣南路1588号

仪器/样品名称:

Description

无局放工频成套试验装置

型号/规格:

Model/Type

HMYDQW-10kVA/100kV

制造厂商:

Manufacturer

武汉国电华美电气设备有限公司

出厂编号:

Serial No.

2203028

管理号:

Asset No.

样品接收日期:

Date of Receipt

2023-03-22

Y M D

结果:

Conclusion

所校准项目合格(Passed at Calibration Items)

校准日期:

Date of Calibration

2023-03-22

Y M D

建议下次校准日期:

Due Date

2024-03-21

Y M D

校准:

Calibration by

程越

审核:

Inspected by

徐继胜

授权签字人:

Approved Signatory

方文潮

证书专用章  
(Stamp)

本实验室地址: 广东省广州市番禺区石碁镇农科所南街8号 @力赛计量实验室

Address: No.8.South Street Shi Ji Institute Guangzhou.Guangdong.China

联系电话(Tel): 020-31134076

投诉电话(CT): 020-31104772

传真(Fax): 020-31134076

邮政编码(Post): 511400

公司网址(Web): www.lisaitest.com

电子邮件(E-mail): cal@lisaitest.com



# 校准说明

## DIRECTIONS OF CALIBRATION

证书编号:  
Certificate No.



2WB23031752543-0007

本机构质量管理体系符合ISO/IEC17025标准要求。

The quality system is in accordance with ISO/IEC17025.

1. 证书内页中"P"代表"合格";"F"代表"不合格";"N/A"代表"不适用"。

In the datasheet,"P" represents "Pass" and "F" represents "Fail" and "N/A" represents "Not Applicable".

2. 本证书编号具有唯一性, 后缀若带有"A~Z"的证书为替换证书, 自发出后原证书即刻作废。

Each certificate has a unique number,The suffix of "A~Z" will be added to the number as a replacement of the old version. The original certificate will be officially invalid once the new certificate number is issued.

3. 环境条件。(Environmental condition during the calibration).

温度(Temperature): 22°C 相对湿度 (Relative Humidity): 58 %

4. 校准地点。(Place of the Calibration).

客户现场(1楼车间)

5. 未加盖“证书报告专用章”无效,报告无校准或检测、审核、授权签字人签章无效。

The report is invalid without the official stamp, report is invalid without 'report stamp', The report is invalid without the signatures of Approval and Reviewer.

6. 样品及生产/研制方的相关信息由委托方提供并由委托方承担相应责任。

The information of the samples and the manufacturer/developer shall be provided by the applicants and applicant shall bear the corresponding responsibilities.

7. 本次校准的技术依据及CNAS认可范围, 超出范围的内容未被认可。详细认可范围请查看CNAS网站中注册编号为L7127的证书附件。(Reference documents and accredited scope by CNAS for the Calibration beyond which isn't accredited. Please see the attachment of certificate No.L7127 on CNAS website for details.)

JJF(力赛)1702-2017 交流高压发生器校准规范 AC High Voltage Generator Calibration Specification

8. 本次校准使用的主要计量标准器具。(Main standards of measurement used in the Calibration).

名称 Description	出厂编号 Serial No.	证书号/有效期 Certificate No./ Due Date	溯源机构 Traceability Institute	技术特征 Technique Character
交直流分压器 AC / DC Voltage Divider	17022403	3WC22050640084-1514 / 2023-05-25	广州力赛计量检测有限公司(LiSai Metrology)	DCV:±0.5% ACV:±1%
数字万用表	19010081	3WC22090614841-3131 / 2023-09-26	广州力赛计量检测有限公司(LiSai Metrology)	DCV:Urel=0.03%, ACV:Urel=0.2% DCI:Urel=0.1%, ACI:Urel=0.5% OHM:Urel=0.05%(k=2)



# 广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO.,LTD

证书编号:  
Certificate No.



中国认可  
国际互认  
校准  
CALIBRATION  
CNAS L7127

2WB23031752543-0007

1、外观以及一般性检查: 正常  
In view of External and Generality check: Pass

## 2、交流高压的校准( Calibration of AC High Voltage):

量程 (Range)	指示值 Indicated	标准值 Reference	误差 Error	不确定度 Uncertainty	允许误差 MPE	结论 Conclusion
(kV)	(kV)	(kV)	(kV)	(%)	(kV)	(Pass/Fail)
100	20.00	20.14	-0.14	1.2	± 1.50	P
	40.00	40.26	-0.26	1.2	± 1.50	P
	60.00	60.42	-0.42	1.2	± 1.50	P
	80.00	79.56	0.44	1.2	± 1.50	P
	100.00	99.52	0.48	1.2	± 1.50	P

附: 关于测量结果不确定度的说明:

appendix: Directions of uncertainty in the calibration

1.依据(Reference document)

JJF 1059.1-2012测量不确定度评定与表示

(JJF 1059.1-2012 Evaluation and Expression of Uncertainty in Measurement)

2.本报告给出的扩展不确定度是由合成标准不确定度乘以包含概率约为95%时对应的包含因子k得到的。

The expanded uncertainty given in this report is obtained by multiplying the combined standard uncertainty by the corresponding coverage factor k when the coverage probability of about 95%.

-----以下空白(The below is blank)-----



广州力赛计量检测有限公司

GUANGZHOU LISAI CALIBRATION AND TESTING CO.,LTD



中国认可  
国际互认  
校准  
CALIBRATION  
CNAS L7127

证书编号:  
Certificate No.



2WB23031752543-0007

本页空白